

# Instruction Sheet

## Rectangular mount for Flo-Tote 3 sensor installation

---

### Safety information

Obey all danger, warning and caution statements. If these statements are not obeyed, there is a risk of serious injury to the operator or damage to the equipment. Do not use or install this equipment in any manner other than that specified in this document.

**WARNING**

**Indicates a potentially or imminently hazardous situation which, if not avoided, could result in death or serious injury.**

**Important Note:** Indicates a situation which, if not avoided, may cause damage to the instrument. Information that requires special emphasis.

### Installation


**WARNING**

**Potential confined space hazards. Training in pre-entry testing, ventilation, entry procedures, evacuation/rescue procedures and safety work practices is necessary before entering confined spaces.**

**WARNING**

**There can be potentially life-threatening hazards associated with the installation of the Flo-Tote 3 sensor. This instruction sheet does not fully address all safety and installation issues and must not be used as a substitute for the Flo-Tote 3 user manual. Refer to the Flo-Tote 3 user manual DOC026.XX.00787.**

**Important Note:** During installation, make sure that there are no spaces between the rectangular mount and the channel wall where debris can collect.




Customer-supplied items:




- 2 anchor bolts that are suitable for the channel wall
- Masonry drill bit
- Tools for the anchor bolt installation
- Neoprene or duct tape to wrap around the sensor cable and mounting plate
- Electrical tape to attach the sensor cable to the rectangular mount





Examine the package for the components that are shown below. The mounting hardware that is supplied with the sensor is also required.

Refer to the illustrations on page 3–4 to assemble the mounting plate and pole, attach the sensor to the mounting plate and install the assembly in the pipe or channel.




75012-XX







5




6



7



8





# Contact information

---

## United States

### U.S.A.

Hach Company  
4539 Metropolitan Court  
Frederick, MD 21704-9452, U.S.A.  
Tel (800) 368-2723  
Fax 301-874-8459  
hachflowsales@hach.com  
www.hachflow.com

### Information required

- Account number (if available)
- Name and phone number
- Purchase order number
- Brief description or model number
- Billing address
- Shipping address
- Catalog number
- Quantity

## European Union

### France

Polymetron/Dr. Lange  
Division Lab  
33 rue du ballon  
F-93165 Noisy le Grand  
Tel.: ++33/(0)1/ 48 15 68 70  
Fax: ++33/(0)1/ 48 15 68 79  
email: info@drlange.fr  
www.drlange.fr

### Great Britain

Dr. Lange (UK) Ltd.  
Lennox Road  
GB-Basingstoke  
Hampshire  
RG22 4AP  
Tel.: ++44/(0)12 56/ 33 34 03  
Fax: ++44/(0)12 56/ 33 07 24  
email: info@drlange.co.uk  
www.drlange.co.uk

### All countries except France and Great Britain

Flow-Tronic  
RUE J.H. COOL 19a  
B-4840 Welkenraedt  
Belgium  
Tel: +-32-87-899797 or 899799  
Fax: +-32-87-899790  
email: sales@mmi-europe.com

## Outside the U.S.A. and EU

Hach Company maintains a worldwide network of dealers and distributors. To find the representative for your location, send an email to hachflowsales@hach.com or visit www.hachflow.com.

## Technical support

Technical and Customer Service Department personnel are eager to answer questions about our products and their use. In the U.S.A., call 800-368-2723. Outside the U.S.A., send an email to hachflowservice@hach.com or call 301-874-5599.